

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00899 Calib. Date: 08/28/2013 Calib. No.: 00020
Certification File No.: 00900 Cert. Date: 08/28/2013 Cert. No.: 00013
Linearity File No.: 00901 Lin. Date: 08/28/2013 Lin. No.: 00013
Solution File No.: 00946 Soln. Date: 10/31/2013 Soln. No.: 00156
Sequential File No.: 00946 File Date: 10/31/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 06/17/2015
Solution Control Lot: 13F114 Bottle No.: 1101

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:42S		
Control 1 EC	0.100%	08:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:43S		
Control 2 EC	0.099%	08:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44S		
Control 3 EC	0.099%	08:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PK #129

Badge No.: 129

Date: 10/31/2013