

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00655	Calib. Date:	08/28/2013
Certification File No.:	00656	Cert. Date:	08/28/2013
Linearity File No.:	00657	Lin. Date:	08/28/2013
Solution File No.:	00681	Soln. Date:	10/31/2013
Sequential File No.:	00681	File Date:	10/31/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	1113

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:43S		
Control 1 EC	0.098%	10:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		
Control 2 EC	0.096%	10:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46S		
Control 3 EC	0.096%	10:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL #129

Badge No.: 129

Date: 10/31/2013