

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00682	Cert. No.: 00011
Sequential File No.:	00682	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 11/08/2013
		Soln. No.: 00136
		File Date: 11/08/2013

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDAB-0014
Control Solution %: 0.100%		Expires: 06/17/2015
Solution Control Lot: 13F114		Bottle No.: 1109

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:56S		
Control 1 EC	0.100%	23:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:58S		
Control 2 EC	0.097%	23:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:59S		
Control 3 EC	0.097%	00:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:00S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CZECH

First Name: TODD

MI: A

Signature: 

Badge No.: 127

Date: 11/09/2013