

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00899	Calib. Date:	08/28/2013
Certification File No.:	00900	Cert. No.:	00020
Linearity File No.:	00901	Cert. Date:	08/28/2013
Solution File No.:	00954	Lin. No.:	00013
Sequential File No.:	00954	Soln. Date:	11/15/2013
		Soln. No.:	00158
		File Date:	11/15/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	1123

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:35S		
Control 1 EC	0.100%	09:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36S		
Control 2 EC	0.099%	09:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37S		
Control 3 EC	0.098%	09:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 11/15/2013