

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00655 Calib. Date: 08/28/2013 Calib. No.: 00013  
Certification File No.: 00656 Cert. Date: 08/28/2013 Cert. No.: 00011  
Linearity File No.: 00657 Lin. Date: 08/28/2013 Lin. No.: 00011  
Solution File No.: 00684 Soln. Date: 11/15/2013 Soln. No.: 00137  
Sequential File No.: 00684 File Date: 11/15/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014  
Control Solution %: 0.100% Expires: 06/17/2015  
Solution Control Lot: 13F114 Bottle No.: 1105

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 09:40S |                |                     |
| Control 1 EC      | 0.100% | 09:41S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 09:41S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:42S |                |                     |
| Control 2 EC      | 0.097% | 09:42S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 09:42S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:43S |                |                     |
| Control 3 EC      | 0.098% | 09:44S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 09:44S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:44S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL K #129

Badge No.: 129

Date: 11/15/2013