

# Calibrating Unit New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037   |
| Location:               | WALL TOWNSHIP POLICE  |                         |
| Calibration File No.:   | 00899                 | Calib. Date: 08/28/2013 |
| Certification File No.: | 00900                 | Calib. No.: 00020       |
| Linearity File No.:     | 00901                 | Cert. Date: 08/28/2013  |
| Solution File No.:      | 00965                 | Cert. No.: 00013        |
| Sequential File No.:    | 00965                 | Lin. Date: 08/28/2013   |
|                         |                       | Lin. No.: 00013         |
|                         |                       | Soln. Date: 11/29/2013  |
|                         |                       | Soln. No.: 00159        |
|                         |                       | File Date: 11/29/2013   |
| <br>                    |                       |                         |
| Calibrating Unit:       | WET                   | Model No.: CU-34        |
| Control Solution %:     | 0.100%                | Serial No.: DDAB-0015   |
| Solution Control Lot:   | 13F114                | Expires: 06/17/2015     |
|                         |                       | Bottle No.: 1118        |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 03:03S        |                               |                           |
| Control 1 EC      | 0.100%         | 03:03S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 03:03S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 03:04S        |                               |                           |
| Control 2 EC      | 0.098%         | 03:05S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.100%         | 03:05S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 03:05S        |                               |                           |
| Control 3 EC      | 0.097%         | 03:06S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 03:06S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 03:07S        |                               |                           |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KELLY

First Name: JASON

MI: R

Signature: 

Badge No.: 166

Date: 11/29/2013