

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00899	Calib. Date: 08/28/2013
Certification File No.:	00900	Calib. No.: 00020
Linearity File No.:	00901	Cert. Date: 08/28/2013
Solution File No.:	00967	Cert. No.: 00013
Sequential File No.:	00967	Lin. Date: 08/28/2013
		Lin. No.: 00013
		Soln. Date: 12/05/2013
		Soln. No.: 00160
		File Date: 12/05/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0015
Control Solution %:	0.100%		Expires: 06/17/2015
Solution Control Lot:	13F114		Bottle No.: 1110

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:16S		
Control 1 EC	-.---%	13:17S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	13:17S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 12/05/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00899	Calib. Date: 08/28/2013
Certification File No.:	00900	Calib. No.: 00020
Linearity File No.:	00901	Cert. Date: 08/28/2013
Solution File No.:	00968	Cert. No.: 00013
Sequential File No.:	00968	Lin. Date: 08/28/2013
		Lin. No.: 00013
		Soln. Date: 12/05/2013
		Soln. No.: 00161
		File Date: 12/05/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0015
Control Solution %:	0.100%		Expires: 06/17/2015
Solution Control Lot:	13F114		Bottle No.: 1110

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:56S		
Control 1 EC	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57S		
Control 2 EC	0.099%	15:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59S		
Control 3 EC	0.099%	15:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL JS #129

Badge No.: 129

Date: 12/05/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00690	Cert. No.: 00011
Sequential File No.:	00690	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 12/05/2013
		Soln. No.: 00138
		File Date: 12/05/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0014
Control Solution %:	0.100%		Expires: 06/17/2015
Solution Control Lot:	13F114		Bottle No.: 1124

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:19S		
Control 1 EC	0.099%	13:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:21S		
Control 2 EC	0.096%	13:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:22S		
Control 3 EC	0.097%	13:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 12/05/2013