

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00655 | Calib. Date: 08/28/2013 |
| Certification File No.: | 00656 | Cert. No.: 00013 |
| Linearity File No.: | 00657 | Cert. No.: 00011 |
| Solution File No.: | 00691 | Lin. No.: 00011 |
| Sequential File No.: | 00691 | Soln. No.: 00139 |
| | | File Date: 12/19/2013 |

| | | |
|-----------------------|--------|-----------------------|
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDAB-0014 |
| Solution Control Lot: | 13F114 | Expires: 06/17/2015 |
| | | Bottle No.: 1119 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:22S | | |
| Control 1 EC | 0.100% | 09:23S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 09:23S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:24S | | |
| Control 2 EC | 0.097% | 09:25S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 09:25S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:25S | | |
| Control 3 EC | 0.098% | 09:26S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 09:26S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:27S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 12/19/2013

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00899 | Calib. Date: 08/28/2013 |
| Certification File No.: | 00900 | Calib. No.: 00020 |
| Linearity File No.: | 00901 | Cert. Date: 08/28/2013 |
| Solution File No.: | 00973 | Cert. No.: 00013 |
| Sequential File No.: | 00973 | Lin. Date: 08/28/2013 |
| | | Lin. No.: 00013 |
| | | Soln. Date: 12/19/2013 |
| | | Soln. No.: 00162 |
| | | File Date: 12/19/2013 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDAB-0015 |
| Solution Control Lot: | 13F114 | Expires: 06/17/2015 |
| | | Bottle No.: 1111 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 09:20S | | |
| Control 1 EC | 0.100% | 09:20S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 09:20S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:21S | | |
| Control 2 EC | 0.099% | 09:22S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 09:22S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:22S | | |
| Control 3 EC | 0.098% | 09:23S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 09:23S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:24S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK #129

Badge No.: 129

Date: 12/19/2013