

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00899	Calib. Date: 08/28/2013
Certification File No.:	00900	Calib. No.: 00020
Linearity File No.:	00901	Cert. Date: 08/28/2013
Solution File No.:	00978	Cert. No.: 00013
Sequential File No.:	00978	Lin. Date: 08/28/2013
		Lin. No.: 00013
		Soln. Date: 12/31/2013
		Soln. No.: 00163
		File Date: 12/31/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	13F114	Expires: 06/17/2015
		Bottle No.: 1103

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:14S		
Control 1 EC	0.100%	20:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15S		
Control 2 EC	0.099%	20:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17S		
Control 3 EC	0.099%	20:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN	First Name: JAMES	MI: R
Signature: 	Badge No.: 162	Date: 12/31/2013

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 Calib. Date: 08/28/2013 Calib. No.: 00013
Certification File No.: 00656 Cert. Date: 08/28/2013 Cert. No.: 00011
Linearity File No.: 00657 Lin. Date: 08/28/2013 Lin. No.: 00011
Solution File No.: 00694 Soln. Date: 12/27/2013 Soln. No.: 00140
Sequential File No.: 00694 File Date: 12/27/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 07/18/2015
Solution Control Lot: 13G115 Bottle No.: 1210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:54S		
Control 1 EC	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		
Control 2 EC	0.096%	01:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:58S		
Control 3 EC	0.097%	01:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	01:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____



Badge No.: 162

Date: 12/27/2013