

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00899 Calib. Date: 08/28/2013 Calib. No.: 00020
Certification File No.: 00900 Cert. Date: 08/28/2013 Cert. No.: 00013
Linearity File No.: 00901 Lin. Date: 08/28/2013 Lin. No.: 00013
Solution File No.: 00981 Soln. Date: 01/03/2014 Soln. No.: 00164
Sequential File No.: 00981 File Date: 01/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 06/17/2015
Solution Control Lot: 13F114 Bottle No.: 1115

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:48S | | |
| Control 1 EC | 0.100% | 12:49S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 12:49S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:50S | | |
| Control 2 EC | 0.098% | 12:50S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 12:50S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:51S | | |
| Control 3 EC | 0.098% | 12:52S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 12:52S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:52S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 01/03/2014

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 Calib. Date: 08/28/2013 Calib. No.: 00013
Certification File No.: 00656 Cert. Date: 08/28/2013 Cert. No.: 00011
Linearity File No.: 00657 Lin. Date: 08/28/2013 Lin. No.: 00011
Solution File No.: 00698 Soln. Date: 01/03/2014 Soln. No.: 00142
Sequential File No.: 00698 File Date: 01/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 07/18/2015
Solution Control Lot: 13G115 Bottle No.: 1215

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 14:09S | | |
| Control 1 EC | 0.097% | 14:09S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 14:09S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:10S | | |
| Control 2 EC | 0.096% | 14:11S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 14:11S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:12S | | |
| Control 3 EC | 0.096% | 14:13S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 14:13S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:13S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K. #129

Badge No.: 129

Date: 01/03/2014

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00655 Calib. Date: 08/28/2013 Calib. No.: 00013
Certification File No.: 00656 Cert. Date: 08/28/2013 Cert. No.: 00011
Linearity File No.: 00657 Lin. Date: 08/28/2013 Lin. No.: 00011
Solution File No.: 00697 Soln. Date: 01/03/2014 Soln. No.: 00141
Sequential File No.: 00697 File Date: 01/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 07/18/2015
Solution Control Lot: 13G115 Bottle No.: 1215

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:44S | | |
| Control 1 EC | 0.097% | 12:45S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 12:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:46S | | |
| Control 2 EC | 0.094% | 12:46S | 34.0°C | CONTROL TEST FAILED |
| Control 2 IR | 0.098% | 12:46S | 34.0°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 12:47S | | |
| Control 3 EC | 0.096% | 12:48S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.096% | 12:48S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:49S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 01/03/2014