

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00655	Calib. Date:	08/28/2013
Certification File No.:	00656	Cert. Date:	08/28/2013
Linearity File No.:	00657	Lin. Date:	08/28/2013
Solution File No.:	00700	Soln. Date:	01/08/2014
Sequential File No.:	00700	File Date:	01/08/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	13G115	Expires:	07/18/2015
		Bottle No.:	1205

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:21S		
Control 1 EC	0.099%	04:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23S		
Control 2 EC	0.096%	04:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:25S		
Control 3 EC	0.097%	04:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: \_\_\_\_\_

*PTC Daniel J. Santoro 105*

Badge No.: 165

Date: 01/08/2014