

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00702	Cert. No.: 00011
Sequential File No.:	00702	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 01/12/2014
		Soln. No.: 00144
		File Date: 01/12/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0014
Control Solution %:	0.100%		Expires: 07/18/2015
Solution Control Lot:	13G115		Bottle No.: 1204

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:37S		
Control 1 EC	0.100%	21:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:39S		
Control 2 EC	0.096%	21:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 3 EC	0.097%	21:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: 

Badge No.: 165

Date: 01/12/2014