

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00899 Calib. Date: 08/28/2013 Calib. No.: 00020  
Certification File No.: 00900 Cert. Date: 08/28/2013 Cert. No.: 00013  
Linearity File No.: 00901 Lin. Date: 08/28/2013 Lin. No.: 00013  
Solution File No.: 00989 Soln. Date: 01/17/2014 Soln. No.: 00165  
Sequential File No.: 00989 File Date: 01/17/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015  
Control Solution %: 0.100% Expires: 06/17/2015  
Solution Control Lot: 13F114 Bottle No.: 1120

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:09S		
Control 1 EC	0.100%	10:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10S		
Control 2 EC	0.098%	10:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12S		
Control 3 EC	0.098%	10:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 01/17/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00704	Cert. No.: 00011
Sequential File No.:	00704	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 01/17/2014
		Soln. No.: 00145
		File Date: 01/17/2014

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1220

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:00S		
Control 1 EC	0.097%	10:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02S		
Control 2 EC	0.096%	10:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04S		
Control 3 EC	0.097%	10:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*P.K. #129*

Badge No.: 129

Date: 01/17/2014