

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00655	Calib. Date: 08/28/2013
Certification File No.:	00656	Calib. No.: 00013
Linearity File No.:	00657	Cert. Date: 08/28/2013
Solution File No.:	00706	Cert. No.: 00011
Sequential File No.:	00706	Lin. Date: 08/28/2013
		Lin. No.: 00011
		Soln. Date: 01/19/2014
		Soln. No.: 00146
		File Date: 01/19/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0014
Control Solution %:	0.100%			Expires:	07/18/2015
Solution Control Lot:	13G115			Bottle No.:	1209

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:56S		
Control 1 EC	0.101%	06:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:57S		
Control 2 EC	0.097%	06:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:59S		
Control 3 EC	0.097%	06:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZECH

First Name: TODD

MI: A

Signature: _____

Sgt. Tom Goh #127

Badge No.: 127

Date: 01/19/2014