

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00709 | Calib. Date: 02/10/2014 |
| Certification File No.: | 00710 | Calib. No.: 00014 |
| Linearity File No.: | 00711 | Cert. Date: 02/10/2014 |
| Solution File No.: | 00714 | Cert. No.: 00012 |
| Sequential File No.: | 00714 | Lin. Date: 02/10/2014 |
| | | Lin. No.: 00012 |
| | | Soln. No.: 00149 |
| | | File Date: 02/12/2014 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXB S3-114 |
| Solution Control Lot: | 13G115 | Expires: 07/18/2015 |
| | | Bottle No.: 1219 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 17:49S | | |
| Control 1 EC | 0.099% | 17:49S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 17:49S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:50S | | |
| Control 2 EC | 0.096% | 17:51S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 17:51S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:52S | | |
| Control 3 EC | 0.097% | 17:53S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 17:53S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:54S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____



Badge No.: 165

Date: 02/12/2014