

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01004 Calib. Date: 02/10/2014 Calib. No.: 00021  
Certification File No.: 01005 Cert. Date: 02/10/2014 Cert. No.: 00014  
Linearity File No.: 01006 Lin. Date: 02/10/2014 Lin. No.: 00014  
Solution File No.: 01012 Soln. Date: 02/16/2014 Soln. No.: 00168  
Sequential File No.: 01012 File Date: 02/16/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 06/17/2015  
Solution Control Lot: 13F114 Bottle No.: 1114

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:53S		
Control 1 EC	0.099%	17:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55S		
Control 2 EC	0.097%	17:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:56S		
Control 3 EC	0.097%	17:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NASH

First Name: STEVEN

MI: P

Signature: \_\_\_\_\_

*[Handwritten Signature]* #157

Badge No.: 157

Date: 02/16/2014

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00709	Calib. Date: 02/10/2014
Certification File No.:	00710	Calib. No.: 00014
Linearity File No.:	00711	Cert. Date: 02/10/2014
Solution File No.:	00717	Cert. No.: 00012
Sequential File No.:	00717	Lin. Date: 02/10/2014
		Lin. No.: 00012
		Soln. Date: 02/16/2014
		Soln. No.: 00150
		File Date: 02/16/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:38S		
Control 1 EC	0.099%	17:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:40S		
Control 2 EC	0.088%	17:40S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.092%	17:40S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:41S		
Control 3 EC	0.090%	17:42S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.088%	17:42S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:43S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: \_\_\_\_\_



Badge No.: 146

Date: 02/16/2014

DEPARTMENT OF  
**Law and Public Safety**  
This is to certify that

**Todd A. Czech**  
**Wall Township Police Department**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C  
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 22nd DAY OF July

TWO THOUSAND AND Eight

  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

ORIGINAL COURSE DATES		
DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>1-5-10</u>	<u>OSPA</u>	<u>Wm Horn</u>
2. <u>1/18/12</u>	<u>OSPA</u>	<u>Adam Standa</u>
3. <u>2-12-14</u>	<u>LAKENURST</u>	<u>Wm Horn</u>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____