

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00709	Calib. Date: 02/10/2014
Certification File No.:	00710	Calib. No.: 00014
Linearity File No.:	00711	Cert. Date: 02/10/2014
Solution File No.:	00718	Cert. No.: 00012
Sequential File No.:	00718	Lin. Date: 02/10/2014
		Lin. No.: 00012
		Soln. Date: 02/20/2014
		Soln. No.: 00151
		File Date: 02/20/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1225

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:45S		
Control 1 EC	0.100%	18:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:47S		
Control 2 EC	0.096%	18:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:48S		
Control 3 EC	0.097%	18:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 02/20/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00709	Calib. Date: 02/10/2014
Certification File No.:	00710	Calib. No.: 00014
Linearity File No.:	00711	Cert. Date: 02/10/2014
Solution File No.:	00719	Cert. No.: 00012
Sequential File No.:	00719	Lin. Date: 02/10/2014
		Lin. No.: 00012
		Soln. Date: 03/02/2014
		Soln. No.: 00152
		File Date: 03/02/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:57S		
Control 1 EC	-.---%	08:58S	22.7°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:58S	22.7°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 03/02/2014

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01004	Calib. Date: 02/10/2014
Certification File No.:	01005	Calib. No.: 00021
Linearity File No.:	01006	Cert. No.: 00014
Solution File No.:	01023	Lin. No.: 00014
Sequential File No.:	01023	Soln. No.: 00170
		File Date: 03/02/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1218

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:28S		
Control 1 EC	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29S		
Control 2 EC	0.097%	09:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		
Control 3 EC	0.097%	09:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 03/02/2014