

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01004 **Calib. Date:** 02/10/2014 **Calib. No.:** 00021
Certification File No.: 01005 **Cert. Date:** 02/10/2014 **Cert. No.:** 00014
Linearity File No.: 01006 **Lin. Date:** 02/10/2014 **Lin. No.:** 00014
Solution File No.: 01021 **Soln. Date:** 03/01/2014 **Soln. No.:** 00169
Sequential File No.: 01021 **File Date:** 03/01/2014

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 07/18/2015
Solution Control Lot: 13G115 **Bottle No.:** 1217

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:57S		
Control 1 EC	0.099%	12:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		
Control 2 EC	0.096%	12:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		
Control 3 EC	0.098%	13:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL K #129

Badge No.: 129

Date: 03/01/2014