

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01004	Calib. Date:	02/10/2014
Certification File No.:	01005	Cert. Date:	02/10/2014
Linearity File No.:	01006	Lin. Date:	02/10/2014
Solution File No.:	01030	Soln. Date:	03/09/2014
Sequential File No.:	01030	File Date:	03/09/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	13G115	Expires:	07/18/2015
		Bottle No.:	1222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:24S		
Control 1 EC	0.099%	01:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:25S		
Control 2 EC	0.097%	01:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:27S		
Control 3 EC	0.097%	01:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZECH

First Name: TODD

MI: A

Signature: Sgt. Todd Czech #127

Badge No.: 127

Date: 03/09/2014