

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01004	Calib. Date: 02/10/2014
Certification File No.:	01005	Calib. No.: 00021
Linearity File No.:	01006	Cert. Date: 02/10/2014
Solution File No.:	01037	Cert. No.: 00014
Sequential File No.:	01037	Lin. Date: 02/10/2014
		Lin. No.: 00014
		Soln. Date: 03/15/2014
		Soln. No.: 00173
		File Date: 03/15/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %:	0.100%		Expires: 07/18/2015
Solution Control Lot:	13G115		Bottle No.: 1220

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:39S		
Control 1 EC	0.098%	08:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41S		
Control 2 EC	0.096%	08:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42S		
Control 3 EC	0.097%	08:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLTER

First Name: STEVEN

MI: R

Signature: \_\_\_\_\_

*SGT. Steven R. Wolter #114*

Badge No.: 114

Date: 03/15/2014

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01004	Calib. Date:	02/10/2014
Certification File No.:	01005	Cert. No.:	00021
Linearity File No.:	01006	Cert. Date:	02/10/2014
Solution File No.:	01036	Lin. No.:	00014
Sequential File No.:	01036	Lin. Date:	02/10/2014
		Soln. Date:	03/14/2014
		Soln. No.:	00172
		File Date:	03/14/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	13G115	Expires:	07/18/2015
		Bottle No.:	1223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:47S		
Control 1 EC	0.096%	18:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	18:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:48S		
Control 2 EC	0.095%	18:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	18:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		
Control 3 EC	0.093%	18:50S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.097%	18:50S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	18:51S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

Badge No.: 162

Date: 03/14/2014