

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01004	Calib. Date:	02/10/2014
Certification File No.:	01005	Cert. Date:	02/10/2014
Linearity File No.:	01006	Lin. Date:	02/10/2014
Solution File No.:	01040	Soln. Date:	03/19/2014
Sequential File No.:	01040	File Date:	03/19/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	07/18/2015
Solution Control Lot:	13G115			Bottle No.:	1216

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.099%	07:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.096%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 3 EC	0.096%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_



Badge No.: 162

Date: 03/19/2014