

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01004	Calib. No.: 00021
Certification File No.:	01005	Cert. No.: 00014
Linearity File No.:	01006	Lin. No.: 00014
Solution File No.:	01048	Soln. No.: 00177
Sequential File No.:	01048	File Date: 03/30/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1213

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:01S		
Control 1 EC	0.099%	02:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:02S		
Control 2 EC	0.098%	02:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:04S		
Control 3 EC	0.098%	02:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	02:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: CADIGAN  First Name: JAMES MI: R
 Signature: _____ Badge No.: 162 Date: 03/30/2014