

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01004	Calib. Date: 02/10/2014
Certification File No.:	01005	Calib. No.: 00021
Linearity File No.:	01006	Cert. Date: 02/10/2014
Solution File No.:	01059	Cert. No.: 00014
Sequential File No.:	01059	Lin. Date: 02/10/2014
		Lin. No.: 00014
		Soln. Date: 04/10/2014
		Soln. No.: 00179
		File Date: 04/10/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 1203

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:07D		
Control 1 EC	0.098%	10:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 2 EC	0.096%	10:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 3 EC	0.098%	10:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL #129*

Badge No.: 129

Date: 04/10/2014