

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01062	Calib. Date: 05/06/2014
Certification File No.:	01063	Calib. No.: 00023
Linearity File No.:	01064	Cert. Date: 05/06/2014
Solution File No.:	01086	Cert. No.: 00015
Sequential File No.:	01086	Lin. Date: 05/06/2014
		Lin. No.: 00015
		Soln. Date: 05/24/2014
		Soln. No.: 00184
		File Date: 05/24/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %:	0.100%		Expires: 09/10/2015
Solution Control Lot:	131119		Bottle No.: 0279

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:03D		
Control 1 EC	0.097%	20:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:04D		
Control 2 EC	0.097%	20:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:05D		
Control 3 EC	0.097%	20:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: _____



Badge No.: 152

Date: 05/24/2014