

# Calibrating Unit New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037   |
| Location:               | WALL TOWNSHIP POLICE  |                         |
| Calibration File No.:   | 01062                 | Calib. Date: 05/06/2014 |
| Certification File No.: | 01063                 | Calib. No.: 00023       |
| Linearity File No.:     | 01064                 | Cert. Date: 05/06/2014  |
| Solution File No.:      | 01085                 | Cert. No.: 00015        |
| Sequential File No.:    | 01085                 | Lin. Date: 05/06/2014   |
|                         |                       | Lin. No.: 00015         |
|                         |                       | Soln. Date: 05/24/2014  |
|                         |                       | Soln. No.: 00183        |
|                         |                       | File Date: 05/24/2014   |
| <br>                    |                       |                         |
| Calibrating Unit:       | WET                   | Model No.: CU-34        |
| Control Solution %:     | 0.100%                | Serial No.: DDXB S3-113 |
| Solution Control Lot:   | 13I119                | Expires: 09/10/2015     |
|                         |                       | Bottle No.: 0285        |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 04:12D        |                               |                           |
| Control 1 EC      | 0.097%         | 04:12D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.097%         | 04:12D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:13D        |                               |                           |
| Control 2 EC      | 0.097%         | 04:13D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 04:13D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:14D        |                               |                           |
| Control 3 EC      | 0.097%         | 04:14D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.097%         | 04:14D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:15D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

Badge No.: 162

Date: 05/24/2014