

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01062	Calib. Date: 05/06/2014
Certification File No.:	01063	Calib. No.: 00023
Linearity File No.:	01064	Cert. Date: 05/06/2014
Solution File No.:	01093	Cert. No.: 00015
Sequential File No.:	01093	Lin. Date: 05/06/2014
		Lin. No.: 00015
		Soln. Date: 05/30/2014
		Soln. No.: 00185
		File Date: 05/30/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	131119	Expires: 09/10/2015
		Bottle No.: 0290

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:51D		
Control 1 EC	0.096%	13:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52D		
Control 2 EC	0.097%	13:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53D		
Control 3 EC	0.096%	13:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

DK #129

Badge No.: 129

Date: 05/30/2014