

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00722	Calib. Date: 04/11/2014
Certification File No.:	00723	Calib. No.: 00016
Linearity File No.:	00724	Cert. Date: 04/11/2014
Solution File No.:	00759	Cert. No.: 00013
Sequential File No.:	00759	Lin. Date: 04/11/2014
		Lin. No.: 00013
		Soln. Date: 05/30/2014
		Soln. No.: 00157
		File Date: 05/30/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	131119	Expires: 09/10/2015
		Bottle No.: 0284

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:57D		
Control 1 EC	0.099%	13:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:58D		
Control 2 EC	0.098%	13:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:59D		
Control 3 EC	0.098%	14:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PK #129*

Badge No.: 129

Date: 05/30/2014