

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01062 Calib. Date: 05/06/2014 Calib. No.: 00023  
Certification File No.: 01063 Cert. Date: 05/06/2014 Cert. No.: 00015  
Linearity File No.: 01064 Lin. Date: 05/06/2014 Lin. No.: 00015  
Solution File No.: 01114 Soln. Date: 06/24/2014 Soln. No.: 00187  
Sequential File No.: 01114 File Date: 06/24/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 0289

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:52D		
Control 1 EC	0.098%	13:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53D		
Control 2 EC	0.097%	13:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		
Control 3 EC	0.097%	13:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: \_\_\_\_\_

*PK #129*

Badge No.: 129

Date: 06/24/2014