

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01062	Calib. Date: 05/06/2014
Certification File No.:	01063	Cert. No.: 00023
Linearity File No.:	01064	Cert. No.: 00015
Solution File No.:	01121	Lin. No.: 00015
Sequential File No.:	01121	Soln. No.: 00188
		File Date: 07/05/2014

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0283

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:34D		
Control 1 EC	0.097%	14:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:36D		
Control 2 EC	0.097%	14:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		
Control 3 EC	0.097%	14:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ERTLE

First Name: ERIK

MI: T

Signature: _____



Badge No.: 105

Date: 07/05/2014