

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01062	Calib. Date:	05/06/2014
Certification File No.:	01063	Cert. No.:	00023
Linearity File No.:	01064	Cert. Date:	05/06/2014
Solution File No.:	01137	Lin. Date:	05/06/2014
Sequential File No.:	01137	Lin. No.:	00015
		Soln. Date:	07/18/2014
		Soln. No.:	00190
		File Date:	07/18/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0294

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:26D		
Control 1 EC	0.097%	22:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		
Control 2 EC	0.097%	22:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28D		
Control 3 EC	0.097%	22:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____



Badge No.: 146

Date: 07/18/2014