

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00722	Calib. Date: 04/11/2014
Certification File No.:	00723	Calib. No.: 00016
Linearity File No.:	00724	Cert. Date: 04/11/2014
Solution File No.:	00779	Cert. No.: 00013
Sequential File No.:	00779	Lin. Date: 04/11/2014
		Lin. No.: 00013
		Soln. Date: 07/18/2014
		Soln. No.: 00159
		File Date: 07/18/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0288

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:19D		
Control 1 EC	0.099%	22:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20D		
Control 2 EC	0.098%	22:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:22D		
Control 3 EC	0.099%	22:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____



Badge No.: 146

Date: 07/18/2014