

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01062	Calib. Date:	05/06/2014
Certification File No.:	01063	Cert. No.:	00023
Linearity File No.:	01064	Cert. Date:	05/06/2014
Solution File No.:	01155	Lin. No.:	00015
Sequential File No.:	01155	Soln. Date:	08/08/2014
		Soln. No.:	00192
		File Date:	08/08/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	131119	Expires:	09/10/2015
		Bottle No.:	0282

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:19D		
Control 1 EC	0.098%	06:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 2 EC	0.097%	06:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 3 EC	0.097%	06:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____

Badge No.: 174

Date: 08/08/2014

*E noted
null 8/14/14*