

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: WALL TOWNSHIP POLICE
 Calibration File No.: 00722
 Certification File No.: 00723
 Linearity File No.: 00724
 Solution File No.: 00791
 Sequential File No.: 00791


Serial No.: ARZL-0146
 Calib. Date: 04/11/2014
 Cert. Date: 04/11/2014
 Lin. Date: 04/11/2014
 Soln. Date: 08/08/2014
 File Date: 08/08/2014
 Calib. No.: 00016
 Cert. No.: 00013
 Lin. No.: 00013
 Soln. No.: 00160

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 131119
 Model No.: CU-34
 Serial No.: DDXB S3-114
 Expires: 09/10/2015
 Bottle No.: 0276

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:33D		
Control 1 EC	0.097%	13:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		
Control 2 EC	0.097%	13:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		
Control 3 EC	0.098%	13:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: KAZANOWSKY
 First Name: PTL.
 MI: A
 Signature: 
 Badge No.: 129
 Date: 08/08/2014