

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01062	Calib. Date: 05/06/2014
Certification File No.:	01063	Calib. No.: 00023
Linearity File No.:	01064	Cert. Date: 05/06/2014
Solution File No.:	01164	Cert. No.: 00015
Sequential File No.:	01164	Lin. Date: 05/06/2014
		Lin. No.: 00015
		Soln. Date: 08/12/2014
		Soln. No.: 00193
		File Date: 08/12/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %:	0.100%		Expires: 09/10/2015
Solution Control Lot:	13I119		Bottle No.: 0298

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:21D		
Control 1 EC	0.098%	04:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:22D		
Control 2 EC	0.098%	04:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23D		
Control 3 EC	0.097%	04:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	04:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: JASON

MI: R

Signature: _____

[Handwritten Signature] #1166

Badge No.: 166

Date: 08/12/2014