

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01062 Calib. Date: 05/06/2014 Calib. No.: 00023
Certification File No.: 01063 Cert. Date: 05/06/2014 Cert. No.: 00015
Linearity File No.: 01064 Lin. Date: 05/06/2014 Lin. No.: 00015
Solution File No.: 01171 Soln. Date: 08/19/2014 Soln. No.: 00194
Sequential File No.: 01171 File Date: 08/19/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 09/10/2015
Solution Control Lot: 13I119 Bottle No.: 0292

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:03D		
Control 1 EC	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:04D		
Control 2 EC	0.096%	00:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		
Control 3 EC	0.096%	00:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	00:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRAMLICH

First Name: JONATHAN

MI: R

Signature: PHI. Jonathan Gramlich 179

Badge No.: 179

Date: 08/19/2014