

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00722	Calib. Date: 04/11/2014
Certification File No.:	00723	Calib. No.: 00016
Linearity File No.:	00724	Cert. Date: 04/11/2014
Solution File No.:	00810	Cert. No.: 00013
Sequential File No.:	00810	Lin. Date: 04/11/2014
		Lin. No.: 00013
		Soln. Date: 08/27/2014
		Soln. No.: 00161
		File Date: 08/27/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	131119	Expires: 09/10/2015
		Bottle No.: 2960

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:56D		
Control 1 EC	0.100%	12:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:57D		
Control 2 EC	0.099%	12:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58D		
Control 3 EC	0.099%	12:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CZECH	First Name: TODD	MI: A
Signature: <u>Sgt Tom Ga 127</u>	Badge No.: 127	Date: 08/27/2014

2-mailed to Lauber
9-3-14