

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01062	Calib. Date: 05/06/2014
Certification File No.:	01063	Calib. No.: 00023
Linearity File No.:	01064	Cert. Date: 05/06/2014
Solution File No.:	01184	Cert. No.: 00015
Sequential File No.:	01184	Lin. Date: 05/06/2014
		Lin. No.: 00015
		Soln. Date: 09/14/2014
		Soln. No.: 00196
		File Date: 09/14/2014

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %: 0.100%		Expires: 09/10/2015
Solution Control Lot: 131119		Bottle No.: 0287

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:49D		
Control 1 EC	0.098%	04:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:50D		
Control 2 EC	0.098%	04:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:52D		
Control 3 EC	0.097%	04:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: \_\_\_\_\_

Badge No.: 152

Date: 09/14/2014