

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01191 Calib. Date: 10/10/2014 Calib. No.: 00024  
Certification File No.: 01192 Cert. Date: 10/10/2014 Cert. No.: 00016  
Linearity File No.: 01193 Lin. Date: 10/10/2014 Lin. No.: 00016  
Solution File No.: 01199 Soln. Date: 11/07/2014 Soln. No.: 00199  
Sequential File No.: 01199 File Date: 11/07/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 0299

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:53S		
Control 1 EC	0.099%	14:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54S		
Control 2 EC	0.099%	14:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55S		
Control 3 EC	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL. K #129*

Badge No.: 129

Date: 11/07/2014