

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00833	Calib. Date:	10/10/2014
Certification File No.:	00834	Cert. Date:	10/10/2014
Linearity File No.:	00835	Lin. Date:	10/10/2014
Solution File No.:	00841	Soln. Date:	11/07/2014
Sequential File No.:	00841	File Date:	11/07/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	0148

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:57S		
Control 1 EC	0.100%	14:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 2 EC	0.100%	14:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		
Control 3 EC	0.100%	15:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

RLC K #129

Badge No.: 129

Date: 11/07/2014