

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00833	Calib. Date:	10/10/2014
Certification File No.:	00834	Cert. No.:	00017
Linearity File No.:	00835	Cert. Date:	10/10/2014
Solution File No.:	00844	Lin. Date:	10/10/2014
Sequential File No.:	00844	Lin. No.:	00014
		Soln. Date:	11/22/2014
		Soln. No.:	00165
		File Date:	11/22/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	14D129	Expires:	04/29/2016
		Bottle No.:	0142

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:30S		
Control 1 EC	0.101%	21:30S	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	21:30S	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	21:31S		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____

Badge No.: 146

Date: 11/22/2014