

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00833	Calib. Date: 10/10/2014
Certification File No.:	00834	Cert. No.: 00017
Linearity File No.:	00835	Cert. Date: 10/10/2014
Solution File No.:	00845	Lin. No.: 00014
Sequential File No.:	00845	Soln. Date: 11/23/2014
		Soln. No.: 00166
		File Date: 11/23/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0149

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:21S		
Control 1 EC	0.100%	02:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23S		
Control 2 EC	0.100%	02:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:24S		
Control 3 EC	0.099%	02:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____



Badge No.: 146

Date: 11/23/2014