

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01191	Calib. Date: 10/10/2014
Certification File No.:	01192	Calib. No.: 00024
Linearity File No.:	01193	Cert. Date: 10/10/2014
Solution File No.:	01214	Cert. No.: 00016
Sequential File No.:	01214	Lin. Date: 10/10/2014
		Lin. No.: 00016
		Soln. Date: 11/26/2014
		Soln. No.: 00201
		File Date: 11/26/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0146

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:37S		
Control 1 EC	0.099%	07:38S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.099%	07:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		
Control 3 EC	0.099%	07:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 11/26/2014