

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00833 Calib. Date: 10/10/2014 Calib. No.: 00017
Certification File No.: 00834 Cert. Date: 10/10/2014 Cert. No.: 00014
Linearity File No.: 00835 Lin. Date: 10/10/2014 Lin. No.: 00014
Solution File No.: 00848 Soln. Date: 11/26/2014 Soln. No.: 00167
Sequential File No.: 00848 File Date: 11/26/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 04/29/2016
Solution Control Lot: 14D129 Bottle No.: 0137

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:44S		
Control 1 EC	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45S		
Control 2 EC	0.098%	07:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		
Control 3 EC	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 11/26/2014