

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01191 Calib. Date: 10/10/2014 Calib. No.: 00024
Certification File No.: 01192 Cert. Date: 10/10/2014 Cert. No.: 00016
Linearity File No.: 01193 Lin. Date: 10/10/2014 Lin. No.: 00016
Solution File No.: 01220 Soln. Date: 12/19/2014 Soln. No.: 00202
Sequential File No.: 01220 File Date: 12/19/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 04/29/2016
Solution Control Lot: 14D129 Bottle No.: 0138

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:03S		
Control 1 EC	0.100%	09:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05S		
Control 2 EC	0.099%	09:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		
Control 3 EC	0.100%	09:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: _____

Badge No.: 152

Date: 12/19/2014