

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00833 Calib. Date: 10/10/2014 Calib. No.: 00017  
Certification File No.: 00834 Cert. Date: 10/10/2014 Cert. No.: 00014  
Linearity File No.: 00835 Lin. Date: 10/10/2014 Lin. No.: 00014  
Solution File No.: 00856 Soln. Date: 12/19/2014 Soln. No.: 00168  
Sequential File No.: 00856 File Date: 12/19/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015  
Control Solution %: 0.100% Expires: 04/29/2016  
Solution Control Lot: 14D129 Bottle No.: 0131

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:09S		
Control 1 EC	0.100%	09:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10S		
Control 2 EC	0.100%	09:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12S		
Control 3 EC	0.100%	09:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: \_\_\_\_\_



Badge No.: 152

Date: 12/19/2014