

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00833	Calib. Date: 10/10/2014
Certification File No.:	00834	Calib. No.: 00017
Linearity File No.:	00835	Cert. Date: 10/10/2014
Solution File No.:	00860	Cert. No.: 00014
Sequential File No.:	00860	Lin. Date: 10/10/2014
		Lin. No.: 00014
		Soln. Date: 12/30/2014
		Soln. No.: 00169
		File Date: 12/30/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0015
Control Solution %:	0.100%		Expires: 04/29/2016
Solution Control Lot:	14D129		Bottle No.: 0126

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:54S		
Control 1 EC	0.099%	11:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55S		
Control 2 EC	0.099%	11:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56S		
Control 3 EC	0.099%	11:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

DK #129

Badge No.: 129

Date: 12/30/2014