

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00833	Calib. Date: 10/10/2014
Certification File No.:	00834	Calib. No.: 00017
Linearity File No.:	00835	Cert. Date: 10/10/2014
Solution File No.:	00869	Cert. No.: 00014
Sequential File No.:	00869	Lin. Date: 10/10/2014
		Lin. No.: 00014
		Soln. Date: 01/28/2015
		Soln. No.: 00170
		File Date: 01/28/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0127

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:52S		
Control 1 EC	0.100%	09:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		
Control 2 EC	0.100%	09:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		
Control 3 EC	0.100%	09:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: \_\_\_\_\_

Badge No.: 152

Date: 01/28/2015