

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01191	Calib. Date: 10/10/2014
Certification File No.:	01192	Calib. No.: 00024
Linearity File No.:	01193	Cert. Date: 10/10/2014
Solution File No.:	01240	Cert. No.: 00016
Sequential File No.:	01240	Lin. Date: 10/10/2014
		Lin. No.: 00016
		Soln. Date: 02/20/2015
		Soln. No.: 00205
		File Date: 02/20/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0132

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:19S		
Control 1 EC	0.101%	17:19S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20S		
Control 2 EC	0.100%	17:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:22S		
Control 3 EC	0.100%	17:22S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

Badge No.: 129

Date: 02/20/2015