

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 01191 | Calib. Date: 10/10/2014 |
| Certification File No.: | 01192 | Calib. No.: 00024 |
| Linearity File No.: | 01193 | Cert. Date: 10/10/2014 |
| Solution File No.: | 01244 | Cert. No.: 00016 |
| Sequential File No.: | 01244 | Lin. Date: 10/10/2014 |
| | | Lin. No.: 00016 |
| | | Soln. Date: 03/01/2015 |
| | | Soln. No.: 00207 |
| | | File Date: 03/01/2015 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDAB-0014 |
| Solution Control Lot: | 14D129 | Expires: 04/29/2016 |
| | | Bottle No.: 0141 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 14:30S | | |
| Control 1 EC | 0.100% | 14:31S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 14:31S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:32S | | |
| Control 2 EC | 0.100% | 14:32S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 14:32S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:33S | | |
| Control 3 EC | 0.101% | 14:34S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 14:34S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:34S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 03/01/2015