

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00833	Calib. Date: 10/10/2014
Certification File No.:	00834	Calib. No.: 00017
Linearity File No.:	00835	Cert. Date: 10/10/2014
Solution File No.:	00878	Cert. No.: 00014
Sequential File No.:	00878	Lin. Date: 10/10/2014
		Lin. No.: 00014
		Soln. Date: 03/01/2015
		Soln. No.: 00173
		File Date: 03/01/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0129

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:47S		
Control 1 EC	0.100%	11:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48S		
Control 2 EC	0.100%	11:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50S		
Control 3 EC	0.100%	11:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK #129

Badge No.: 129

Date: 03/01/2015